

Calculating F_{DO_2} for Mixtures of Air and Oxygen

The need to combine air and oxygen frequently arises during anesthesia and critical care. Primarily, the nitrogenous component of air functions to reduce the concentration of oxygen and to decrease the risk of oxygen toxicity.¹ This dilution is also mandatory during laser surgery because high concentrations of oxygen support combustion.²⁻⁴ Use of 100% oxygen also can result in absorption atelectasis.⁵ The oxygen concentration delivered (F_{DO_2}), when air and oxygen are mixed, is calculated by Equation 1

$$F_{DO_2} = \frac{(1.0)(\dot{V}_{O_2}) + (0.21)(\dot{V}_{air})}{\dot{V}_{O_2} + \dot{V}_{air}}$$

where \dot{V}_{O_2} = O_2 flowrate and
 \dot{V}_{air} = air flowrate.

If the flowrates of air and O_2 are known, the F_{DO_2} can be quickly calculated by Equation 2

$$F_{DO_2} = (0.79) \left(\frac{\dot{V}_{O_2}}{\dot{V}_{total}} \right) + 0.21,$$

where \dot{V}_{total} = the total flowrate
($\dot{V}_{air} + \dot{V}_{O_2}$).

As shown in the Appendix, Equation 2 is just a simplified form of Equation 1. It is especially practical for rapid estimates and often can be used without a calculator. All flowrates must be in the same units (ordinarily $mL \cdot \min^{-1}$ or $L \cdot \min^{-1}$). F_{DO_2} is the fraction of the total gas flow that is oxygen. The coefficients of the equation, 0.79 and 0.21, correspond to the percentage of nitrogen and oxygen in air, respectively. Equation 2 can be easily rearranged

to solve for \dot{V}_{O_2} . Example: With an air flowrate of $4.5 L \cdot \min^{-1}$ and oxygen flow of $0.5 L \cdot \min^{-1}$ (total flow, or $\dot{V}_{total} = 5 L \cdot \min^{-1}$), the resulting F_{DO_2} is

$$0.79 \left(\frac{0.5}{5} \right) + 0.21 = 0.29,$$

ie, the oxygen fraction.

An estimate can be obtained by rounding off 0.79 to 0.80 and 0.21 to 0.20 and, in this example $0.5/5 = 0.1$ and thus

$$(0.80)(0.1) + 0.20 = 0.28,$$

ie, the oxygen fraction.

By making use of simple arithmetic shortcuts, the reader can make this method work expeditiously. By Equation 1, it may be difficult to determine \dot{V}_{O_2} for a given F_{DO_2} , but Equation 2 can be easily manipulated to solve for \dot{V}_{O_2} .

$$\dot{V}_{O_2} = \frac{(\dot{V}_{total})(F_{DO_2} - 0.21)}{0.79}$$

Example: For an F_{DO_2} of 0.3 (30% O_2), with \dot{V}_{total} of $8 L \cdot \min^{-1}$, the flowrate for O_2 will be

$$\frac{(8)(0.3 - 0.21)}{0.79} = 0.91 L \cdot \min^{-1}.$$

Again, rounding off and knowing that $8/0.8 = 10$ and $0.3 - 0.21 = 0.09$, the solution can be estimated

$$\left(\frac{8}{0.8} \right) (0.3 - 0.21) = 0.9 L \cdot \min^{-1}.$$

\dot{V}_{air} can be determined by subtracting the flowrate of O_2 from the \dot{V}_{total} ($\dot{V}_{air} = \dot{V}_{total} - \dot{V}_{O_2}$). In the preceding example, the flowrate for air would be $8.0 - 0.91 = 7.09 L \cdot \min^{-1}$. Use of these equations should reduce the 'trial and error' (and anxiety!) associated with setting air and oxygen flowrates.

Glen Atlas MD MSc

Resident

Department of Anesthesiology

SUNY Health Science Center

Syracuse, New York

REFERENCES

1. Klein J. Normobaric pulmonary oxygen toxicity. *Anesth Analg* 1990;70:195-207.
2. Macintosh R, Mushin WW, Epstein HG. Deflagrations and detonations in fuel mixtures. *Physics for the anaesthetist*, 3rd edition. Oxford: Blackwell Scientific Publications, 1963:326-357.
3. Wolf GL, Simpson JL. Flammability of endotracheal tubes in oxygen and nitrous oxide enriched atmospheres. *Anesthesiology* 1987;67:236-239.
4. Schramm VL Jr, Mattox DE, Stool SE. Acute management of laser-ignited intratracheal explosion. *Laryngoscope* 1989;91(9, Part 1):1417-1426.
5. Nunn JF. *Applied respiratory physiology*, 3rd edition. London: Butterworth & Co, 1987:442-443.

APPENDIX

Equation 2 can be derived from Equation 1 by simple algebra. First, \dot{V}_{O_2} can be expressed as the sum of two fractions $(0.79)(\dot{V}_{O_2})$ and $(0.21)(\dot{V}_{O_2})$.

$$F_{DO_2} = \frac{(1.0)(\dot{V}_{O_2}) + (0.21)(\dot{V}_{air})}{\dot{V}_{O_2} + \dot{V}_{air}} = \frac{(0.79 + 0.21)(\dot{V}_{O_2}) + (0.21)(\dot{V}_{air})}{\dot{V}_{O_2} + \dot{V}_{air}}$$

Because both \dot{V}_{O_2} and \dot{V}_{air} are multiplied by a factor of 0.21, they can be collected. The term $[\dot{V}_{O_2} + \dot{V}_{air}]$ then is eliminated from both numerator and denominator of that part of the equation:

$$F_{DO_2} = \frac{(0.79)(V_{O_2})}{V_{O_2} + V_{air}} + \frac{(0.21)(V_{O_2} + V_{air})}{V_{O_2} + V_{air}}$$

$$= \frac{(0.79)(V_{O_2})}{V_{O_2} + V_{air}} + 0.21$$

Replacing $[\dot{V}_{O_2} + \dot{V}_{air}]$ with the equation $\dot{V}_{total} = \dot{V}_{O_2} + \dot{V}_{air}$ yields the simplified result:

$$F_{DO_2} = (0.79) \left(\frac{V_{O_2}}{V_{total}} \right) + 0.21.$$

Professional Literacy Revisited

Last Fall Semester (1991), my students and I read with interest Robert Weilacher's editorial in RESPIRATORY CARE entitled "Professional Literacy."¹ I feel it professionally incumbent upon me to add some comments.

I have taught two advanced-level Respiratory Therapist reading courses (REST 2525 and REST 2526) in our Registry Curriculum every year for the past 19 years. In those courses, my students and I have read and evaluated practically every article published in RESPIRATORY CARE since 1972. I want to reassure my fellow Texan from Palestine that at least one academic educator not only uses but relies on RESPIRATORY CARE in his work!

I (and my students) have never seen the quality of articles in the Journal as good as they are at present. We are proud to be a part of a profession from whose peers such high quality, original, and scholarly work as that published in RESPIRATORY CARE this past year can be generated. The recent academic quality reflected in our monthly science journal should make each member of the Journal's Editorial Board proud—especially because it makes AARC members and respiratory care students proud.

I totally agree with Mr Weilacher that the Journal is *the* one tool always available to me and my stu-

dents in that "constant struggle to remain professionally contemporary."¹ Permit me to illustrate: This Spring Semester (1992), one of the topics emphasized in my course is mechanical ventilation. What a quality synthesis of up-to-date information awaited us in the Journal papers by such internationally acclaimed authors as Richard Branson RRT, Robert Chatburn RRT, Roger Goldberg MD, Dean Hess MEd RRT, Robert Kacmarek PhD RRT, and David Pierson MD—just to name a few! In Volume 36 (1991), topics related to mechanical ventilation ranged from ventilator performance during hyperbaric compression to bronchodilator administration, circuit compression volume, and patient work of breathing, and numbered more than 35—not including letters! My students and I found the Chatburn paper² on ventilator classification to be the single, most helpful paper published in the past 10 years.

I think that sometimes we take our Journal too much for granted; too often we do not take the time to pat on the back those who continuously produce a high quality product year after year. I have certainly failed to do so in times past, and my students and I simply want to express our thanks to the Editorial Board and the Journal's authors for adding so much to our professional enrichment and pride.

Ralph E Bartel MEd RCP RRT
Department Head
Respiratory Therapist (REST)
Program
Houston Community College

1. Weilacher R. Professional literacy (editorial). *Respir Care* 1991;36:1083-1084.
2. Chatburn R L. A new system for understanding mechanical ventilators. *Respir Care* 1991;36:872-874.

For your convenience, and direct access, the advertisers in this issue and their phone numbers are listed below. Please use this directory for requesting written material or for any question you may have.

HELP LINES

Arabian Medical.....	800-537-1026
Burroughs Wellcome Co	919-248-3000
Bear Medical Systems Inc (Ventilator Upgrade)	800-366-2327
Bear Medical Systems Inc (Infant Ventilator)	800-BEAR-MED
Ciba-Corning	800-255-3232
Criticare Systems	800-458-4615
HealthScan Products	800-962-1266
Instrumentation Laboratory	800-955-9525
Kenneth Norris Cancer Hospital	213-324-2660
Life Design Systems Inc	800-872-9010
Medical Graphics Corp	800-950-5597
MSA Catalyst Research	800-851-4500
PPG Industries Inc	619-552-9022
Puritan-Bennett	800-255-6773
RFB Technologies	407-367-8611
Ross Laboratories	614-227-3189
Sherwood Medical	800-325-7472
Siemens Medical Systems, Patient Care Division	800-944-9046
WB Saunders Company	800-545-2522

LAST CALL FOR OPEN FORUM ABSTRACTS!

It's time to submit your abstracts for possible presentation at the OPEN FORUM during the AARC Annual Meeting in San Antonio, Texas, December 12-15, 1992.

Deadline for submission is June 6.

See Page 484 of this issue
for more information.